



DECLARATION

Labbox Labware declares under the manufacturer responsibility that this good:

MCEN-00B; MCEN-00G

is in conformity with the:

Directive 2014/30/EU - EMC (European Electromagnetic Compatibility

Directive 2014/35/EU - LVD (Low Voltage Directive)

Directive 2011/65/EU and 2015/863/EU - RoHS (Restriction of Hazardous Substances)

Moreover, this equipment has tested against the standards:

EN 61010-1 Safety Requirements for Electrical Equipment for Measurement

Control, and Laboratory Use – Part 1: General Requirements

EN 61326-1 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

EN 61000-3-2 Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions

EN 61000-3-3 Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker

EN 61010-2-020 Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges

This document has been produced electronically and it is valid without signature: 09/10/2024